

## **Magnetic Resonance Imaging (MRI) Compatibility**

The following product has been evaluated for safety in Magnetic Resonance Environments.

Dansac Nova™ 1 Infant Pouch (SKU 818-00 and 818-10)

The results of the evaluation are as follows:

**Dansac Nova™ 1 Infant Pouch (SKUs 818-00 and 818-10):** This product, with the soft wire closure attached, has been evaluated for safety in the Magnetic Resonance Environment. The product is labeled as “MR Conditional.” A patient with this device can be scanned safely with this device in an MR system under the following conditions:

- a. Static magnetic field of 1.5-Tesla and 3-Tesla, only
- b. Maximum spatial gradient magnetic field of 466-Gauss/cm (4.7-T/m)
- c. Maximum MR system reported, whole body averaged specific absorption rate (SAR) of 2-W/kg for 15 minutes of scanning (i.e.: per pulse sequence) in the Normal Operating Mode
- d. A 15.2cm (6 inch) X 2.5cm (1 inch) length of adhesive tape must be applied over the metallic strip to help stabilize this product when it is necessary for a patient to undergo an MRI examination at 1.5-Tesla or 3-Tesla

Prior to use, be sure to read the Instructions for Use for information regarding Intended Use, Contraindications, Warnings, Precautions, and Instructions.

Dansac, the Dansac logo, and Nova are trademarks of Dansac A/S. © 2025 Dansac A/S.

